

IgA Serological Response for the Diagnosis of Mycobacterium abscessus Infections in Patients with Cystic Fibrosis

Vincent Le Moigne, Anne-Laure Roux, Hélène Mahoudo, Gaëtan Christien, Agnès Ferroni, Oana Dumitrescu, Gérard Lina, Jean-Philippe Bouchara,

Patrick Plésiat, Jean-Louis Gaillard, et al.

▶ To cite this version:

Vincent Le Moigne, Anne-Laure Roux, Hélène Mahoudo, Gaëtan Christien, Agnès Ferroni, et al.. IgA Serological Response for the Diagnosis of Mycobacterium abscessus Infections in Patients with Cystic Fibrosis. Microbiology Spectrum, 2022, 10 (3), pp.00192-22. 10.1128/spectrum.00192-22. hal-03682405

HAL Id: hal-03682405 https://uvsq.hal.science/hal-03682405v1

Submitted on 29 Jun2022

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution 4.0 International License





IgA Serological Response for the Diagnosis of *Mycobacterium abscessus* Infections in Patients with Cystic Fibrosis

[®] Vincent Le Moigne,^a Anne-Laure Roux,^{a,b} Hélène Mahoudo,^a Gaëtan Christien,^a Agnès Ferroni,^c Oana Dumitrescu,^{d,e} [®] Gérard Lina,^{d,e} Jean-Philippe Bouchara,^f [®] Patrick Plésiat,^g Jean-Louis Gaillard,^{a,b} Stéphane Canaan,^h [®] Geneviève Héry-Arnaud,^{i,j} Jean-Louis Herrmann^{a,k}

^aUniversité Paris Saclay, UVSQ, Inserm, Infection et Inflammation, Montigny-le-Bretonneux, France

^bAP-HP, GHU Paris Saclay, Hôpital Ambroise Paré, Service de Microbiologie, Boulogne-Billancourt, France

cAP-HP, GHU Paris, Hôpital Necker-Enfants Malades, Service de Microbiologie, Paris, France

^dHospices Civils de Lyon, Hôpital de la Croix Rousse-Centre de Biologie Nord, Institut des Agents Infectieux, Laboratoire de Bactériologie, Lyon, France

«Centre International de Recherche en Infectiologie, INSERM U1111, Université de Lyon, Lyon, France

fUNIV Angers, UNIV Brest, Groupe d'Etude des Interactions Hôte-Pathogène (GEIHP, EA 3142), SFR 4208 ICAT, CHU, Angers, France

eLaboratoire de Bactériologie, CHRU de Besançon, UMR CNRS 6249 Chrono-Environnement, Faculté de Médecine-Pharmacie, Université de Bourgogne Franche-Comté, Besançon, France

hAix-Marseille Univ, CNRS, LISM, IMM FR3479, Marseille, France

ⁱDépartement de Bactériologie-Virologie, Hygiène et Parasitologie-Mycologie, Centre Hospitalier Universitaire (CHU) de Brest, Brest, France

ilnserm, EFS, UMR 1078, France; Génétique, Génomique fonctionnelle et Biotechnologies », GGB, Brest, France

^kAP-HP, GHU Paris Saclay, Hôpital Raymond Poincaré, Service de Microbiologie, Garches, France

ABSTRACT The immunoglobulin A (IgA) status of cystic fibrosis (CF) patients, presenting with or without a non-tuberculous mycobacterial (NTM) infection, has to date not been fully elucidated toward two antigenic preparations previously described. We have chosen to determine the clinical values of an IgA ELISA for the diagnosis of NTM and/or Mycobacterium abscessus infections in CF patients. One hundred and 73 sera from CF patients, comprising 33 patients with M. abscessus positive cultures, and 31 non-CF healthy controls were assessed. IgA levels were evaluated by indirect ELISAs using a surface antigenic extract named TLR2eF for TLR2 positive extract and a recombinant protein, the phospholipase C (rMAB_0555 or rPLC). These assays revealed a sensitivity of 52.6% (95% CI = 35.8% to 69%) and 42.1% (95% CI = 26.3% to 59.2%) using TLR2eF and rPLC, respectively, and respective specificities of 92.6% (95% CI = 87.5% to 96.1%) and 92% (95% CI = 86.7% to 95.7%) for samples culture positive for *M. abscessus*. Overall sensitivity and specificity of 66.7% and 85.4%, respectively, were calculated for IgA detection in *M. abscessus*-culture positive CF patients, when we combine the results of the two used antigens, thus demonstrating the efficiency in detection of positive cases for these two antigens with IgA isotype. CF patients with a positive culture for M. abscessus had the highest IgA titers against TLR2eF and rPLC. The diagnosis of NTM infections, including those due to M. abscessus, can be improved by the addition of an IgA serological assay, especially when cultures, for example, are negative. Based on these promising results, a serological follow-up of a larger number of patients should be performed to determine if the IgA response may be correlated with an active/acute infection state or a very recent infection.

IMPORTANCE *Mycobacterium abscessus* is currently the most frequently isolated rapid growing *mycobacterium* in human pathology and the major one involved in lung infections. It has recently emerged as responsible for severe pulmonary infections in patients with cystic fibrosis (CF) or those who have undergone lung transplantation. In addition, it represents the most antibiotic resistant mycobacterial species. However, despite its increasing clinical importance, very little is known about the use of *M. abscessus* parietal compounds and the host response. This has led to the development of

Editor Silvia T. Cardona, University of Manitoba Copyright © 2022 Le Moigne et al. This is an open-access article distributed under the terms of the Creative Commons Attribution 4.0 International license.

Address correspondence to Vincent Le Moigne, vincent.le-moigne@uvsq.fr, or Jean-Louis Herrmann, jean.louis-herrmann@aphp.fr. The authors declare no conflict of interest.

Received 20 January 2022 Accepted 30 March 2022 Published 18 May 2022 serological tests to measure the antibody response in infected patients, and potentially to link this to the culture of respiratory samples. Herein, we describe an important analysis of the serological IgA response from CF patients, and we demonstrate the full diagnostic usefulness of this assay in the diagnosis of NTM infections, and more particularly *M. abscessus*, in CF patients.

KEYWORDS non-tuberculous mycobacteria, cystic fibrosis, IgA, serology, serodiagnosis, ELISA

gA are antibodies classically produced by mucous membranes, but with regard to mycobacterial infections, much attention has been devoted to the evaluation of specific IgG in most serological studies. However, IgA antibodies have been shown to be useful in the diagnosis of various infections due to fungi (*Candida albicans, Trichosporon cutaneum*) or parasites (*Leishmania, Toxoplasma*), as well as viruses, including SARS-CoV-2 (1, 2) and bacteria. Among bacterial infections, the IgA response has been investigated in infections caused by *Bordetella pertussis* (3), *Chlamydia trachomatis*, and mycobacteria, and especially *Mycobacterium tuberculosis*, with about three quarters of all IgA studies concerning this pathogen within the *Mycobacterium* genus (4–8). Other mycobacterial infections targeted in IgA serological studies, especially in cystic fibrosis (CF) patients, were *Mycobacterium avium* (9) or *M. abscessus* (10, 11).

An enzyme immunoassay kit developed by the group of Kitada (12, 13) was used by Jeong et al. (10) to detect IgA antibodies reacting to a glycopeptidolipid (GPL) core antigen derived from *M. avium* complex. With this reagent, they were able to detect patients infected with *M. abscessus* or *Mycobacterium massiliense* without differentiating them from other mycobacteria belonging to the avium complex, whereas they differentiated those with *M. tuberculosis*, linked to the fact that GPLs exist only in non-tuberculous mycobacterials (NTMs). The same reagent (Tauns Laboratories) was used in three recent studies in 2020 (14–16) to identify patients infected with *M. abscessus*. Although, this assay did not allow the differentiation of sera from patients infected with *M. abscessus* complex mycobacteria or with the *M. avium* species complex (MAC), it predicted a favorable outcome, revealing a decrease over time of antibody level in surgically treated patients (15). Decrease of GPL core IgA antibodies in patient sera was also correlated with antibiotic treatment efficacy (15), indicative of the usefulness of IgA serum level to predict the recurrence of disease or treatment efficacy.

As part of a diagnostic accuracy study, we demonstrate the importance of IgG antibodies detection for the diagnosis of NTM infections in CF patients, and more accurately for those infected by *M. abscessus* infections with two antigenic preparations (17). Using the same serum collection and the same antigens, we evaluated the potential of the IgA detection technique in enhancing the diagnosis of NTM infections, in order to decipher a more complete humoral response in CF patients infected by NTM, and to see if the IgA response might be a new diagnostic tool for detecting NTMinfected CF patients, and more specifically *M. abscessus*-infected CF patients.

RESULTS

We first evaluated the diagnostic value of the recombinant PLC (rPLC) and surfaceextract TLR2 activators-enriched fraction (TLR2eF) IgA ELISAs by analyzing the sera of 59 CF patients NTM-infected to the sera of 114 CF patients NTM-non-infected or a control group (31 healthy controls [HC] non-CF). The IgA antibody titers against TLR2eF were significantly different between both CF-groups (P < 0.0001) (Fig. 1A). The area under the receiver operating characteristic (ROC) curve (AUC) was 0.645, allowing the determination of a cut-off value of 0.751 (Fig. S1A). This values allowed us to obtain specificity and sensitivity of 92.5% (95% CI = 86.9% to 96.2%) and 38.2% (95% CI = 25.4% to 52.3%), respectively, for the TLR2eF ELISA for detecting NTM culture-positive individuals.

Similarly, IgA antibody titers against rPLC distinctly differentiated both groups (P < 0.0001) (Fig. 1B). The area under the ROC curve (AUC) was 0.655, allowing the determination of a cut-




FIG 1 IgA response of NTM positive culture, non-NTM positive culture groups and healthy control (HC) group using TLR2eF (A) and rPLC (B) as antigens. Each dot represents one patient in the scattergrams. Horizontal lines represent the mean and vertical bars SDs. Values are presented in Tables S1A and S1B in the supplemental material for each antigenic sample, respectively. Chosen cutoff values (test positivity threshold) are respectively 0.751 (A) and 1.274 (B) (dotted horizontal lines). P < 0.001 for comparisons of NTM-culture positive groups *versus* the non-NTM group plus the HC group for TLR2eF and rPLC. Beneath each figure is represented the corresponding 2 by 2 table for NTM-culture positive patients. n.s., nonsignificant; *, P < 0.05; **, P < 0.01; ***, P < 0.001; ****, P < 0.001.

off value of 1.274 (Fig. S1B). This value allowed us to obtain specificity and sensitivity of 91.8% (95% CI = 86.1% to 95.7%) and 30.9% (95% CI = 19.1% to 44.8%), respectively, specifically for the detection of NTM-culture positive-CF patients with the rPLC ELISA slightly lower than those obtained with TLR2eF. Positive and negative predictive values (PPV and NPV), positive and negative likelihood ratios (PLR and NLR), and accuracy were calculated based on a prevalence of 3.6% according to the results obtained previously (18). We provide the values for both antigens in Tables S1A and S1B. The most interesting score was obtained with TLR2eF as antigenic extract.

Then, with a between mycobacterial species comparison, the TLR2eF antigenic extract allows to obtain Optical density (OD) values that significantly separated CF patients with *M. abscessus* (Mabs) positive cultures from the other patient groups (P < 0.01) (Fig. 2A), except for the CF patient group with *M. intracellulare* (Mint) positive cultures (P = 0.25), when separated from the group of CF patients with *M. avium* complex (MAC) positive cultures (Fig. S3B). When Mabs culture positive-CF patients were compared to the other patient groups including the healthy controls, the specificity and sensitivity for the detection of Mabs culture positive-CF patients were 92.6% (95% CI = 87.5% to 96.1%) and 52.6% (95% CI = 35.8% to 69%), respectively. The specificity and sensitivity for detecting MAC culture positive-CF patients were by comparison 92% and 9.5%, respectively (Table S2B).

We also demonstrated distinct and significant differences between CF groups when we used rPLC as antigenic target, with a higher average response for Mabs culture positive-CF patients compared with all other patient groups (P < 0.01) with the exception of MAC culture positive-CF patients (P = 0.32) or other NTM culture positive-CF patients (P = 0.15) (Fig. 2B). The specificity and sensitivity for the detection of Mabs culture positive-CF patients were 92% (95% CI = 86.7% to 95.7%) and 42.1% (95% CI: = 26.3% to 59.2%), respectively. Specificity and sensitivity for detecting MAC culture positive-CF patients were, by comparison, 92% and 19.1%, respectively (Table S2B). Although the




FIG 2 IgA response of the different cystic fibrosis (CF) patient groups and healthy control (HC) group using TLR2eF (A) or rPLC (B) as antigen. Each dot represents one patient in the scattergrams. In MAC group, yellow squares represent *M. avium*-, pink squares represent *M. chimaera*- and red squares *M. intracellulare*-culture positive-CF patients. Horizontal lines represent the mean and vertical bars SDs. Chosen cut-off values (test positivity threshold) are 0.751 and 1.274 in A and B, respectively (dotted horizontal lines). Beneath each figure is represented the corresponding 2 by 2 table, for Mabs culture positive-CF patients, and see the 2 by 2 Tables S2A and S2B in the supplemental material for MAC culture positive-CF patients. n.s., nonsignificant; *, P < 0.05; **, P < 0.01; ****, P < 0.001; ****, P < 0.001.

differences were not significant, the OD values measured for Mint culture positive-CF patients were higher than those obtained for *M. avium* plus *M. chimaera* (Mav/Mchm) culture positive-CF patients (P = 0.085) and, to a lesser extent, for Mabs culture positive-CF patients (P = 0.65) (Fig. S3B). PPV and NPV, PLR and NLR, and accuracy for both antigens are given in Tables S3A and S3B. Here again, the most convincing performances were found with TLR2eF antigenic extract.

Finally, we put together results for each antigen (TLR2eF and rPLC), and classified sera as a positive result when at least one of the two assays was positive, and in a negative category when the results were negative for each of the two ELISAs. A comparison of these data with culture results was then achieved (Table 1 and 2). The specificity and sensitivity to detect CF patients with NTM positive cultures, calculated from Table 1, were 85.5% and 44.1%, respectively, and 85.4% and 66.7%, respectively, calculated from Table 2, when focusing on patients with Mabs positive cultures. PPV and NPV, PLR and NLR, and accuracy for these two tables are presented just below their corresponding two by two table from which they were calculated, with a NPV of 98.6% when comparing patients with Mabs positive cultures.

TABLE 1 Two by two table when combining TLR2eF and rPLC serology results versus NTM culture^a

TLR2eF or rPLC	NTM positive culture	NTM negative culture	Total
Positive serology	26	21	47
Negative serology	33	124	157
Total	59	145	204
Parameter	Valu	e	95% Cl
PPV (%)	10.2		6.5 to 15.6
NPV (%)	97.6		97.0 to 98.1
PLR	3.0		1.9 to 5.0
NLR	0.65		0.52 to 0.83
Accuracy (%)	84		78.3 to 88.8

^aCorresponding values of positive predictive value (PPV), negative predictive value (NPV), positive likelihood ratio (PLR), negative likelihood ratio (NLR), and accuracy and their respective confidence interval at 95% (95% Cl).

TABLE 2 Two by two table when combining TLR2eF and rPLC serology results versus Mabs culture^{*a*}

TLR2eF or rPLC	Mabs positive culture	Mabs negative culture	Total
Positive serology	22	25	47
Negative serology	11	146	157
Total	33	171	204
Parameter	Value		95% Cl
PPV (%)	14.55		9.9 to 20.8
NPV (%)	98.6		97.7 to 99.1
PLR	4.6		2.95 to 7.05
NLR	0.39		0.24 to 0.64
Accuracy (%)	84.7		79.0 to 89.3

^aCorresponding values of positive predictive value (PPV), negative predictive value (NPV), positive likelihood ratio (PLR), negative likelihood ratio (NLR), and accuracy and their respective confidence interval at 95% (95% CI).

DISCUSSION

The importance of the IgA response against mycobacterial pathogens has been demonstrated in several studies (6–10, 19). Both mucosal and systemic IgA have protective effects and can trigger pro-inflammatory response (20, 21). Furthermore, with a half-life four time shorter than IgG, serum specific IgA might reflect a recent active infection. For these reasons, the evaluation of IgA, alongside IgG measurement, makes sense. In addition, the combination of IgA and IgG responses may help to improve serodiagnosis test accuracy for active NTM and/or *M. abscessus* infection.

In this study, framed as an exploration of the suitability of the serum IgA assay for the diagnosis of NTM infections, we have shown low sensitivity values when each antigen was evaluated separately, around 30% and up to 50% when considering Mabs-positive CFpatients; but with a doubling in sensitivity when the two antigens were combined, culminating in a value of 60%, specifically for the diagnosis of Mabs infections. These results confirm that the use of a cocktail of antigens instead of single antigens increases the sensitivity of mycobacterial serodiagnosis. This might be due to the differential expression of certain antigens during the development of mycobacteria (22). In addition, this multi-antigen analysis gives even better results when IgG and IgA levels are cumulated, like improving sensitivity from 23% to 62% as described by Juliàn et al. (23) for *M. tuberculosis*. Introducing the IgG results obtained from the same cohort (17) allowed us to see that when at least three results are positive, we obtained a sensitivity of 60.6%, a specificity of 95.3% and positive and negative predictive values of 32.6% and 98.5%, respectively (Table S4). Similarly, these sensitivity values can reach 90% if at least one of the tests is positive i.e., either IgA with TLR2e, or IgA with rPLC and likewise IgG (Table S4). Our study has shown that this enzyme immunoassay measuring the IgA response against M. abscessus rPLC and the TLR2eF extract might well complement the IgG data and help to distinguish M. abscessus from M. avium pulmonary infections as well as other infections or lung diseases. Finally and most interestingly, in all cases the negative predictive value is greater than or equal to 97.5%, making a strong argument in the setting of a negative IgG and or IgA response for excluding NTM infection in CF patients. In the context of pharmacological regimes which restore CFTR, or with patients with non-productive cough, such a result might avoid additional investigations or even unnecessary treatment in these patients.

Unlike the only test measuring IgA to detect Mabs infection published to date (10, 14–16), our IgA ELISA was able to differentiate Mabs from Mav/Mchm-infections, but not from Mint-infections although the number of patients was small. Our assay is based on the recognition of Mabs proteins or extracts while the older test is based on detection of antibodies recognizing the GPL core antigen of *M. avium*.

Both ELISAs are currently being tested in a prospective study to determine the prevalence of NTM infections in CF patients (clinical trial no. ID RCB: 2017-A00025-48). In addition, variations in IgA levels over time might be indicative of disease development in the setting of rising levels, or in the case of a decrease, might provide proof of successful antibiotic treatment as it has been shown with TB infection (8). This is also an area of interest which we are studying in this cohort, in addition to the comparative lgA/lgG kinetics.

MATERIALS AND METHODS

Patients and antigens. Serum samples for antibody determination, antigens, or preparations used in this study as well as the distribution of the different groups depending on the results of sputum cultures are the same as those previously described (17).

ELISAs. ELISAs were developed as described previously (17). First, the coating of plates was completed overnight at 4°C with 1 µg/mL of each antigen, recombinant PLC (rPLC), or the surface-extract TLR2 activators-enriched fraction (TLR2eF) diluted in 100 μ L of carbonate-bicarbonate buffer (0.1 M, pH 9.6). Then, plates were washed with phosphate-buffered saline (PBS) twice and blocked by incubation for 1 h at 37°C with 200 µL of PBS containing 1% bovine serum albumin (PBS-BSA). Sera were diluted at a 1/400th dilution in PBS-Tween 20 (0.05% vol/vol) (PBS-T) containing 0.5% BSA (PBS-T-BSA) and added to the plates. Plates were incubated for 90 min at 37°C and, following four washes with PBS-T, alkaline phosphatase-conjugated goat antihuman IgA (Southern biotechnology, Birmingham, USA) diluted in PBS-T-BSA was added. A second incubation of 90 min of the plates was performed at 37°C and four additional washes with PBS-T are realized before the addition of alkaline phosphatase substrate, 100 µL of 1 mg/mL of p-nitrophenylphosphate (Sigma, Saint Quentin Fallavier, France) diluted in diethanolamine buffer (pH 9.8). Plates were finally incubated in the dark at room temperature for 120 min. Ultimately, the spectrophotometric reading of the plates was determined at the wavelength of 405 nm. All serum were tested in duplicate. OD mean \pm standard deviation was calculated for each patient group. The test positivity thresholds, sensitivities, and specificities of the ELISAs were assessed by calculating the AUC ROC curve. ROC curves were generated using GraphPad Prism 6.0 software (GraphPad Software, La Jolla, CA, USA). The closest point to the upper left side corner was used to determine the cut-off value. ROC curves corresponding to Fig. 1 and 2 are presented in Fig. S1 and S2 in the supplemental material.

Data analysis. A Student's *t* test was used for all analyses using GraphPad Prism 6.0 software (GraphPad Software, La Jolla, CA, USA). A two-sided *P* of < 0.05 was considered to be statistically significant.

SUPPLEMENTAL MATERIAL

Supplemental material is available online only. **SUPPLEMENTAL FILE 1**, PDF file, 0.7 MB.

ACKNOWLEDGMENTS

We sincerely thank the members of the MucoMicrobes group (Association Vaincre la Mucoviscidose) for their support. We warmly thank Ben Marshall (University Hospitals of Southampton, United Kingdom) for a thorough review and correction of the manuscript.

This work was supported by grants from the Association Vaincre la Mucoviscidose (RF20110600446/1/3/130).

All authors declare no conflicts of interest regarding the study.

REFERENCES

- Jalkanen P, Pasternack A, Maljanen S, Melén K, Kolehmainen P, Huttunen M, Lundberg R, Tripathi L, Khan H, Ritvos MA, Naves R, Haveri A, Österlund P, Kuivanen S, Jääskeläinen AJ, Kurkela S, Lappalainen M, Rantasärkkä K, Vuorinen T, Hytönen J, Waris M, Tauriainen S, Ritvos O, Kakkola L, Julkunen I. 2021. A combination of N and S antigens with IgA and IgG measurement strengthens the accuracy of SARS-CoV-2 serodiagnostics. J Infect Dis 224: 218–228. https://doi.org/10.1093/infdis/jiab222.
- Jamiruddin MR, Haq MA, Tomizawa K, Kobatake E, Mie M, Ahmed S, Khandker SS, Ali T, Jahan N, Oishee MJ, Khondoker MU, Sil BK, Haque M, Adnan N. 2021. Longitudinal antibody dynamics against structural proteins of SARS-CoV-2 in three COVID-19 patients shows concurrent development of IgA, IgM, and IgG. J Inflamm Res 14:2497–2506. https://doi .org/10.2147/JIR.S313188.
- Fumimoto R, Otsuka N, Kamiya H, Sunagawa T, Tanaka-Taya K, Kamachi K, Shibayama K. 2019. Seroprevalence of IgA and IgM antibodies to *Bordetella pertussis* in healthy Japanese donors: assessment for the serological diagnosis of pertussis. PLoS One 14:e0219255. https://doi.org/10.1371/ journal.pone.0219255.
- Reljic R, Williams A, Ivanyi J. 2006. Mucosal immunotherapy of tuberculosis: is there a value in passive IgA? Tuberculosis (Edinb) 86:179–190. https://doi.org/10.1016/j.tube.2006.01.011.
- Araujo LS, Maciel R. d M, Moraes RM, Trajman A, Saad MHF. 2010. Assessment of the IgA immunoassay diagnostic potential of the *Mycobacterium tuberculo*sis MT10.3-MPT64 fusion protein in tuberculous pleural fluid. Clin Vaccine Immunol 17:1963–1969. https://doi.org/10.1128/CVI.00372-10.

- Le Moigne V, Le Moigne D, Mahana W. 2013. Antibody response to *Mycobacterium tuberculosis* p27-PPE36 antigen in sera of pulmonary tuberculosis patients. Tuberculosis (Edinb) 93:189–191. https://doi.org/10.1016/j.tube.2012.10.006.
- Legesse M, Ameni G, Medhin G, Mamo G, Franken KL, Ottenhoff TH, Bjune G, Abebe F. 2013. IgA response to ESAT-6/CFP-10 and Rv2031 antigens varies in patients with culture-confirmed pulmonary tuberculosis, healthy *Mycobacterium tuberculosis*-infected and non-infected individuals in a tuberculosis endemic setting. Scand J Immunol 78:266–274. https://doi.org/ 10.1111/sji.12080.
- Awoniyi DO, Baumann R, Chegou NN, Kriel B, Jacobs R, Kidd M, Loxton AG, Kaempfer S, Singh M, Walzl G. 2017. Detection of a combination of serum IgG and IgA antibodies against selected mycobacterial targets provides promising diagnostic signatures for active TB. Oncotarget 8: 37525–37537. https://doi.org/10.18632/oncotarget.16401.
- Jhun BW, Kim SY, Park HY, Jeon K, Shin SJ, Koh WJ. 2017. Changes in serum IgA antibody levels against the glycopeptidolipid core antigen during antibiotic treatment of *Mycobacterium avium* complex lung disease. Jpn J Infect Dis 70:582–585. https://doi.org/10.7883/yoken.JJID.2016.523.
- Jeong BH, Kim SY, Jeon K, Lee SY, Shin SJ, Koh WJ. 2013. Serodiagnosis of *Mycobacterium avium* complex and *Mycobacterium abscessus* complex pulmonary disease by use of IgA antibodies to glycopeptidolipid core antigen. J Clin Microbiol 51:2747–2749. https://doi.org/10.1128/JCM.00702-13.
- Nithichanon A, Samer W, Chetchotisakd P, Kewcharoenwong C, Ato M, Lertmemongkolchai G. 2020. Evaluation of plasma anti-GPL-core IgA and IgG

for diagnosis of disseminated non-tuberculous mycobacteria infection. PLoS One 15:e0242598. https://doi.org/10.1371/journal.pone.0242598.


- Kitada S, Maekura R, Toyoshima N, Fujiwara N, Yano I, Ogura T, Ito M, Kobayashi K. 2002. Serodiagnosis of pulmonary disease due to *Mycobacterium avium* complex with an enzyme immunoassay that uses a mixture of glycopeptidolipid antigens. Clin Infect Dis 35:1328–1335. https://doi .org/10.1086/344277.
- Kitada S. 2019. Application of a commercial serodiagnostic kit that measures the serum anti-glycopeptidolipid core IgA antibody in *Mycobacterium avium* complex pulmonary disease. Respir Invest 57:410–414. https://doi.org/10.1016/j.resinv.2019.04.002.
- Kobayashi T, Tsuyuguchi K, Yoshida S, Kurahara Y, Minomo S, Inoue Y, Suzuki K. 2020. Serum immunoglobulin A antibodies to glycopeptidolipid core antigen for Mycobacteroides abscessus complex lung disease. Int J Mycobacteriol 9:76–82.
- 15. Fukushima K, Miki M, Matsumoto Y, Uda E, Yamamoto Y, Kogita Y, Kagawa Y, Matsuki T, Kagawa H, Oshitani Y, Motooka D, Tsujino K, Yoshimura K, Miki K, Hayashi A, Nakamura S, Kitada S, Takeuchi Y, Kida H. 2020. The impact of adjuvant surgical treatment of nontuberculous mycobacterial pulmonary disease on prognosis and outcome. Respir Res 21:153.
- Tam WO, Wong CF, Wong SSY, Kwan CLY. 2020. Serological diagnosis of *Mycobacterium avium* complex lung diseases by enzyme immunoassay of IgA antibodies against MAC-specific glycopeptidolipid core antigen. J Clin Tuberc Other Mycobact Dis 21:100189. https://doi.org/10.1016/j.jctube.2020 .100189.
- 17. Le Moigne V, Roux AL, Mahoudo H, Christien G, Ferroni A, Dumitrescu O, Lina G, Bouchara JP, Plésiat P, Gaillard JL, Canaan S, Héry-Arnaud G,

Herrmann JL, S1569-1993(21)01360–6. 2021. Serological biomarkers for the diagnosis of *Mycobacterium abscessus* infections in cystic fibrosis patients. J Cyst Fibros 21:353–360.

- Roux A-L, Catherinot E, Ripoll F, Soismier N, Macheras E, Ravilly S, Bellis G, Vibet MA, Le Roux E, Lemonnier L, Gutierrez C, Vincent V, Fauroux B, Rottman M, Guillemot D, Gaillard J-L, Herrmann J-L for the OMA Group. 2009. Multicenter study of prevalence of nontuberculous mycobacteria in patients with cystic fibrosis in France. J Clin Microbiol 47:4124–4128. https://doi.org/10.1128/JCM.01257-09.
- Baumann R, Kaempfer S, Chegou NN, Oehlmann W, Spallek R, Loxton AG, van Helden PD, Black GF, Singh M, Walzl G. 2015. A subgroup of latently *Mycobacterium tuberculosis* infected individuals is characterized by consistently elevated IgA responses to several mycobacterial antigens. Mediators Inflamm 2015:1–10. https://doi.org/10.1155/2015/364758.
- 20. Kerr MA. 1990. The structure and function of human IgA. Biochem J 271: 285–296. https://doi.org/10.1042/bj2710285.
- 21. Monteiro RC, Van De Winkel JG. 2003. IgA Fc receptors. Annu Rev Immunol 21:177–204. https://doi.org/10.1146/annurev.immunol.21.120601.141011.
- 22. Jackett PS, Bothamley GH, Batra HV, Mistry A, Young DB, Ivanyi J. 1988. Specificity of antibodies to immunodominant mycobacterial antigens in pulmonary tuberculosis. J Clin Microbiol 26:2313–2318. https://doi.org/10 .1128/jcm.26.11.2313-2318.1988.
- Julián E, Matas L, Alcaide J, Luquin M. 2004. Comparison of antibody responses to a potential combination of specific glycolipids and proteins for test sensitivity improvement in tuberculosis serodiagnosis. Clin Diagn Lab Immunol 11:70–76. https://doi.org/10.1128/cdli.11.1.70-76.2004.

Supplementary Tables S1A and S1B: positive predictive values (PPV), negative predictive values (NPV), positive likelihood ratio (PLR), negative likelihood ratio (NLR) and accuracy and their respective confidence interval at 95% (95% CI) of results presented in Fig. 1.

Α			B		
TLR2eF		95% CI	rPLC		95% CI
PPV (%)	14.9	8.3 to 25.4	PPV (%)	11.5	6.2 to 20.3
NPV (%)	97.5	96.9 to 97.9	NPV (%)	97.2	96.7 to 97.6
PLR	4.7	2.4 to 9.1	PLR	3.5	1.77 to 6.83
NLR	0.7	0.57 to 0.85	NLR	0.78	0.66 to 0.92
Accuracy (%)	90.3	85.5 to 94.0	Accuracy (%)	89.5	84.4 to 93.3




Supplementary Figure S1A and S1B: Sensitivity of the enzyme immunoassays presented in Fig. 1 between NTM-infected and Non-NTM-infected groups plus the control group for the enzyme immunoassays against: (A) the surface-extract TLR2 activators-enriched fraction and (B) PLC recombinant protein.


Receiver operating characteristic (ROC) curve for detection of IgA anti-TLR2eF (A) and anti-rPLC (B) have respectively an area under the curve [AUC] of 0.645 and 0.655.

Supplementary Tables S2A and S2B: Two by two tables showing serology results compared to MAC-positive or negative cultures, using the different antigens: A: TLR2eF ; B: rPLC. Patients with Mabs-positive culture were excluded.

Α				В			
TLR2eF	MAC-	MAC-	Total	rPLC	MAC-	MAC-	Total
	positive	negative			positive	negative	
	culture	culture			culture	culture	
Positive	2	12	14	Positive	4	12	16
serology				serology			
Negative	19	138	157	Negative	17	138	155
serology				serology			
Total	21	150	171	Total	21	150	171




Supplementary Figure S2A and S2B: Sensitivity of the enzyme immunoassays presented in Fig. 2 between Mabs-infected plus Mint-infected groups and other CF patients infected groups plus control group for the enzyme immunoassays against: (A) the surface-extract TLR2 activators-enriched fraction and (B) PLC recombinant protein. Receiver operating characteristic (ROC) curve for detection of IgG anti- TLR2eF (A), and anti-rPLC (B) have respectively an area under the curve [AUC] of 0.758 and 0.711.



Supplementary Figure S3A and S3B: IgA response of the different cystic fibrosis (CF) patient groups with positive cultures for *M. abscessus* (Mabs), *M. intracellulare* (Mint) or *M.* avium plus M. chimaera (Mav/Mchm) opposite to: (A) TLR2eF or the surface-extract TLR2 activating-enriched fraction and (B) rPLC. Each dot represents one patient in the scattergrams. Yellow squares represent M. avium- and pink squares represent M. chimaerapositive cultures in the Mav/Mchm group. Horizontal lines represent the mean and vertical bars SDs.

Supplementary Tables S3A and S3B: positive predictive values (PPV), negative predictive values (NPV), positive likelihood ratio (PLR), negative likelihood ratio (NLR) and accuracy and their respective confidence interval at 95% (95% CI) of results presented in Supplementary Figure 3.

Α			B			
TLR2eF		95% CI		rPLC		95% CI
PPV (%)	19.9	12.1 to 31.0	_	PPV (%)	13.6	7.7 to 22.8
NPV (%)	98.2	97.4 to 98.75		NPV (%)	97.6	96.8 to 98.1
PLR	6.6	3.7 to 12.0		PLR	4.2	2.2 to 7.9
NLR	0.5	0.34 to 0.72		NLR	0.67	0.51 to 0.88
Accuracy (%)	90.47	85.6 to 94.1	_	Accuracy (%)	88.8	83.6 to 92.8



Supplementary Figure S4: IgA and IgG compilation. Results responses previously obtained for IgG antibodies (1) were taken back and added to IgA results: a heat-map/matrix was designed with results obtained for each individual sera with both antigenic preparation, PLC and TLR2eF, and for both serotype. If a sera is above the calculated threshold, this positive result induce a red square. A negative result is represented by a green square. Groups designed according bacterial culture positivity or not are represented on the right.

Supplementary Table S4: IgA and IgG responses compilation. Results previously obtained for IgG antibodies (1) were taken back and added to IgA results: a heat-map/matrix was designed with results obtained for each individual sera with both antigenic preparation, PLC and TLR2eF, and for both serotype (Fig. S4 in the supplemental material). The resulting Table allow to see the number of case in different groups if at least one test is positive, if at least two tests, at least three tests or if the four tests are positive. Corresponding values of sensitivity, specificity, positive predictive value (PPV), negative predictive value (NPV), positive likelihood ratio (PLR), negative likelihood ratio (NLR) and accuracy and their respective confidence interval at 95% (95% CI) are indicated.

	Postitive test number					
	0	At least 1	At least 2	At least 3	4	
Mabs positive culture	2	31	24	20	11	
M. intracellulare	1	4	4	3	2	
MAC (without <i>M. intacellulare</i>)	13	3	1	1	0	
Other NTM	3	2	0	0	0	
Non NTM + HC	108	37	8	4	0	
Total	127	77	37	28	13	

...

	Value (95% CI)					
Consistivity	93.9%	72.7%	60.6%	33.3%		
Sensitivity	(79.8% to 99.3%)	(54.5% to 86.7%)	(42.1% to 77.1%)	(17.9% to 51.8%)		
Specificity	73.1%	92.4%	95.3%	98.8%		
Specificity	(65.8% to 79.6%)	(87.3% to 95.9%)	(91.0% to 98.0%)	(95.8% to 99.9%)		
PPV	11.5%	26.3%	32.6%	51.6%		
PPV	(9.1% to 14.5%)	(16.9% to 38.5%)	(18.9% to 50.1%)	(19.8% to 82.1%)		
NPV	99.7%	98.9%	98.5%	97.5%		
NPV	(98.8% to 99.9%)	(98.1% to 99.4%)	(97.7% to 99.0%)	(96.9% to 98.1%)		
Accuracy	73.8%	91.7%	94.1%	96.5%		
Accuracy	(67.2% to 79.7%)	(87.0% to 95.1%)	(89.9% to 96.9%)	(92.9% to 98.5%)		
PLR	3.5	9.6	12.9	28.5		
FLN	(2.7 to 4.5)	(5.4 to 16.8)	(6.2 to 26.9)	(6.6 to 122.7)		
NLR	0.08	0.30	0.41	0.67		
INER	(0.02 to 0.32)	(0.17 to 0.52)	(0.27 to 0.63)	(0.53 to 0.86)		

REFERENCES

 Le Moigne V, Roux AL, Mahoudo H, Christien G, Ferroni A, Dumitrescu O, Lina G, Bouchara JP, Plésiat P, Gaillard JL, Canaan S, Héry-Arnaud G, Herrmann JL. 2021. Serological biomarkers for the diagnosis of *Mycobacterium abscessus* infections in cystic fibrosis patients. J Cyst Fibros 9:S1569-1993(21)01360-6. doi: 10.1016/j.jcf.2021.08.019.